

COMMITTEE NEW PRODUCT EVALUATION COMMITTEE		DATE: MARCH 01, 2006		
MEMBERS PRESENT: CAROLYN ANCRUM - CHERYL BRIAN - JANET BYRNE – M.E. CANADAY- HELEN FOX – BERNADETTE GARRY – POLLY GUFFIN – GAIL HALE – M.M. McLAIN - SIMEON MOULTRIE – K. JORGENSEN – M.MILLER- KAREN SAVAGE – JUNE STOVALL - S. RIBBLE SANDRA SENTERFEIT – P. SRINIVAS – D. SMITH - BRIAN VEST- KAY JOHNSON-PAM COLLINS- J. DeRHODO Members Absent: C. Alcott, P. Anthony, N. Black, A. Bright, K. Chase, L. Dvorsak, R. Flynn, J. Foster, H. Kokko, P. Malpas, S. Muckenfuss, T. Niksch, P. Riley, B. Roche, W. Rogers, J.Sanderfur, D. Smith, C. Younker, Celeste Phillips, Mary Santanna, J. Stovall Excused: Polly Guffin, & Julia Forrest Guest: PAULA RUKINS & ROBERT WILLIAMS		Meeting March 01, 2006 Time Called to Order: 1:00 p.m. Time Adjourned: 2:00 p.m. Residing Officer: Maria C. Dorger Recording Secretary: Maria C. Dorger Location: 420J, CSB		
AGENDA ITEM	DISCUSSION/CONCLUSIONS	RECOMMENDATION/ACTIONS WHAT/WHEN	WHO	STATUS REVIEW
Minutes from March 01, 2006	Approved		All	Closed
Saf-T Blood Culture Device By: Smith-Medical	Products are much sturdier and has a better connection. Saf-T Blood Cx Device with male Luer (96004) Saf-T Blood Cx. with female Luer (96005) These new products will replace the old Kendall products.	The hospital will deplete the current stock of Kendall product ESI # 2384 and ESI # 2479	Cheryl Brian	Closed
Sage Q-Care Oral Cleansing	Compliance vs reduction in the number of V.A.P. Number of samples for the CCU study. Price of new produce vs. cost of current product.	Compliance on another unit. Track compliance for utilization, before changing out to more costly item.	Joanne Naylor, CCU	Open
Mask by Cardinal Duck Bill Mask	Duck Bill Mask approved in O.R., L&D, Infection Control evaluated. Clinicians fearful that Central may send the white Cardinal “Duck Bill” mask instead of the N-95 and their health could be compromised.	Mask was not approved at this time. Much confusion on the “Duck Bill” mask / N-95, and the new mask. Choose another name for the new mask.	Maria Dorger	Open
Prep Kit	The need to have three different CVL kits will be considered at a later time. Evaluating the cost of new kit vs. cost of betadine/chloraprep. Number of infections since the induction of the Biopatch.	1. Kit with chloraprep –No BioPatch 2. Kit with chloraprep & BioPatch # 364 \$11.60 3. Kit with betadine No chloraprep No BioPatch ESI 4. #1484 \$4.10 Neuro Patients	Maria Dorger	Open

BioPatch for hemapheresis catheters	Requests for the BioPatch to be used for our Dialysis Catheters, 8W, 6E ICU's Hemapheresis, out patient cancer clinics. Order # 3152	Use the BioPatch antimicrobial dressing. Size 7.00 mm vs. the 4.0mm BioPatch.	Laura Milligan	Closed
BioPatch Pediatric and adult PICC lines	Requested the ¾ “ disk with 1.5mm center hole to be used with the neonatal and adult population for their PICC Order # 3151	The use of the BioPatch antimicrobial dressing will reduce the number of infections in this population	Shelia Scarbrough	Closed
WeePee Diapers	Neonatal < 800 grams need a smaller diaper. C.H. currently doesn't carry a diaper that properly fits this population.	Committee approved the use of a Wee Pee Diaper made for population less than 800 grams.	Kathy Jorgensen	Closed
Peds Electrode SilverSircuit	Replace the current electrodes that do not stick. (KittyKat Electrodes) Current practice is changing several times per day.	Replace 25% of the current electrodes with the Silver-Sircuit. 4 cases per month Product # EP00575	Kathy Jorgensen	Closed
Neonatal Electrode PuppyDog	Replace the KittyKat Electrodes that do not stick. Electrodes need to be replaced several times per day.	Replace 75% of the current electrode with the PuppyDog Electrode 11 cases per month Product # 31424785	Kathy Jorgensen	Closed

NEXT PRODUCT EVALUATION MEETING May 03, 2006 @ 1 P.M. ROOM 420J CSB