

**MINUTES**

<b>COMMITTEE: Products Evaluation Committee</b>		<b>DATE: JUNE 7, 2006</b>		
<b>MEMBERS PRESENT:</b> C. ANCRUM, P. ANTHONY, N. BLACK, C. BRIAN, M.E. CANADAY, P. COLLINS, R. FLYNN, J. FORREST, H. FOX, G. HALE, M. MCLAIN, S. RIBBLE, J. MINICK, K. SAVAGE, D. SMITH, P. SRINIVAS,  <b>MEMBERS ABSENT:</b> C. ALCOTT, A. BRIGHT, J. BYRNE, K. CHASE, J. DERHODO, L. DVORSAK, J. FOSTER, B. GARRY, K. JOHNSON, K. JORGESEN, H. KOKKO, P. MALPAS, M. MILLER, S. MOULTRIE, S. MUCKENFUSS, T. NIKSCH, C. PHILLIPS, W. ROGERS, J. SANDERFUR, M. SANTANNA, S. SENTERFEIT, J. STOVALL, B. VEST,		<b>TIME CALLED TO ORDER: 1:00 PM</b>  <b>TIME ADJOURNED: 2:06PM</b>  <b>RESIDING OFFICER: M. DORGER</b>  <b>RECORDING SECRETARY: S. MOULTRIE</b>  <b>LOCATION: 420J, CSB</b>		
<b>AGENDA ITEM</b>	<b>DISCUSSION/CONCLUSIONS</b>	<b>RECOMMENDATION/ACTIONS WHAT/WHEN</b>	<b>WHO</b>	<b>STATUS REVIEW</b>
Huber Needles	Results from surveys are completed. The Gripper Plus will be deleted from hospital use. Too many user errors, resulting in sticks. 82.5% voted on the Huber Plus. Three different needles were available.	Committee recommended Huber plus safety needle to be standardized throughout hospital. Rep will be contacted for inservice.	Maria Dorger	Closed
Chlorprep Swabsticks	B. Rhoton requested chlorprep swabsticks for neonatal, PICU, Pediatric Care and Nursing. Product will be stocked separately in the different area	Committee agreed that swapsticks be stocked separately in the different areas.	Beth Rhoton	Closed
Edward Flo Trac Sensor	Janet Byrne requested flo trac sensor that monitors cardiac output.	Committee recommended to contact rep and further evaluate the flo trac sensor and will implement after approval from Critical Care Ed. Board.	Janet Byrne	Open
Liko Lift Sling	Bernadette Garry requested the Liko Lift Sling for the 6,7 and 8 <sup>th</sup> floors. Sling will help with back injury among the nurse population. Sling can only be used one per patient.	Committee agreed to bring the product in and stock in Central Supply. Further investigation will determine if the product will be a patient charge.	Bernadette Garry	Closed

**Next Products Evaluation Committee Meeting is scheduled for July 5, in Room 420J CSB**